

Temperature control and seating

1. Reducing temperature at the seat-patient interface in carved foam seating

Sarah Greasley, Oxford Centre for Enablement

Additional author: Rick Houghton, Oxford Centre for Enablement

Summary

Preliminary experiments show that open-cell foam has passive cooling benefits in comparison to CM50 foam (closed-cell). Additionally, the use of integrated fans can provide significant cooling, with a more uniform temperature distribution. There is potential to expand this work to direct the air flow, and allow user operated temperature control.

Aims & Objectives

This work aims to reduce seating surface temperature in carved foam seating, both for comfort and pressure ulcer management. A viable testing method for analysing temperature at the seat-patient interface will be established. The passive cooling effects of open-cell foam will be tested, with the option to actively ventilate using fans. The possibility of channelling the fan ventilation to specific areas of higher temperature will also be investigated.

Background

User feedback has indicated that the greatest need for improvement in carved foam seating is the reduction of perspiration by the regulation of interface temperature [1]. It has been found that a subject's definition of a comfortable temperature varies with their current body temperature. Therefore the most comfortable chair should have a neutral effect on body temperature, allowing for air flow to/from the surroundings [2]. It is also known that the severity of pressure injuries correlates with an increase in applied temperature [3]. A reduction in seat-patient interface temperature would therefore have direct clinical benefits. There is currently very limited literature on seat-patient interface temperature in carved foam seating and, as there is no standardised testing method or data collection technique, it is difficult to compare results from different studies. While some studies use complex systems to model subjects, others use either patients or healthy volunteers. Many are also limited in terms of length of data collection or number of subjects. It has been seen in previous studies that the material properties of the cushion has the most significant effect on seat-subject interface temperature [4].

This study looks at the use of a new open-cell foam to make carved foam seating. The foam was successfully machined using a CNC milling machine, and pressure mapping results showed good pressure relieving properties. Temperature sensors (Analog Devices TMP37FT9Z, +5 → +100 °C) were permanently attached to a breathable cover beneath the groin, coccyx, right IT and right mid-thigh so recordings could be taken repeatedly from specific areas of interest. A data logger was used to record temperature variation at each location over a specified time period. Multiple preliminary tests were conducted on a healthy subject under 3 conditions: A) Custom contoured cushion manufactured from standard CM50 foam, B) Custom contoured cushion manufactured from open-cell foam, C) Custom contoured cushion manufactured from open-cell foam with fans providing forced ventilation.

In condition B, the open-cell foam was encased in an ABS shell with ventilation holes to allow for passive air flow. However, in condition C the ABS shell was sealed to direct air flow from the fans towards the patient. The interface temperature was found to vary slightly between repeated tests; however this did not correlate with varying room temperature, which was measured throughout each test and observed to have no effect on results. It was found that in tests A and B, the groin was

at the highest temperature with the same temperature distribution observed for all repeated tests. The open-cell foam provided consistent passive cooling at all 4 specified locations, with temperatures after 45 minutes being an average of 0.9 ± 0.2 °C cooler. Forced ventilation provided increased cooling with a maximum of a 4.0°C of cooling observed at the sacrum after 45 minutes. Cooler temperatures were also seen at all other locations, and statistical analysis showed these decreases to be statistically significant ($p \leq 0.05$) at all locations. Subjective observations from multiple subjects also revealed this cooling to be significant.

Discussion

From these preliminary results, a repeatable testing protocol will be established, and multiple healthy subjects will be tested under each condition for a specified time period. It will be necessary to determine after what time the temperatures plateau for each condition, as this was not observed to happen within the timescales specified by previous research [5]. However, it is expected that the temperature profile will be consistent across all subjects [6]. Only one previous study has used fans to attempt to actively cool patients [1]. This was done by cutting channels for air flow into the foam and resulted in a very uneven temperature distribution. However, the results from our experiment show that active cooling in open-cell foam can produce a far more uniform temperature distribution with a comfortable drop in temperature. Alternatively the air flow could be directed by sealing areas of the open-cell foam or use of non-breathable covers. This will be investigated in the next stage of this research. Additionally the fans were shown to lower the temperature when turned on at any point during the experiment. Cooling could therefore be controlled automatically by switching the fans on/off at specified temperatures or as desired by the user. This will also be investigated.

The main challenge associated with previous use of porous materials in seating is bacterial growth. The open-cell cushions can be washed through under a shower; however, the cushion will have to be dried and these issues will have to be addressed to enable clinical use.

References

1. Scott Chalmers, John Colvin 'Seating Surface Temperature Reduction in Foam Carve Seating,' WestMARC, NHS
2. Herman Miller 'The Attributes of Thermal Comfort: Ergonomic Criteria for the Design of "Breathable" Work Chairs'
3. G.P. Nicholson et al. 'A method for determining the heat transfer and water vapour permeability of patient support systems', Medical Engineering & Physics 21 (1999) 701–712
4. FERGUSON-PELL et al. 'Thermodynamic rigid cushion loading indenter', JRRD 2009
5. Liu Z et al. 'Microenvironment temperature prediction between body and seat interface using autoregressive data-driven model' J Tissue Viability 2015
6. Zhuofu Liu et al. 'Studying thermal characteristics of seating materials by recording temperature from 3 positions at the seat-subject interface' J Tissue Viability 2011

Email: sarah.greasley@ouh.nhs.uk

2. Enhancing comfort and function through temperature controlled back supports - a case study review

Matthew Eveleigh, Sunrise Medical

Summary

Temperature dysregulation can negatively affect both function and comfort for wheelchair users with high level spinal cord injury (SCI), multiple sclerosis (MS) and cerebral palsy (CP). Case study reviews

show the JAY J3 temperature controlled back support enables users to control their own back support temperature, leading to improved comfort and function.

Aims & Objectives

- 1) To clarify the importance of promoting a good microclimate in relation to the back support surface;
- 2) To highlight the negative impact of temperature dysregulation on function and comfort for individuals with high level SCI, MS or CP;
- 3) Demonstrate by case study analysis the use of a temperature controlled J3 back support to improve function and comfort for a SCI, MS or CP sufferer.

Background

Reducing both heat and humidity at the skin surface interface are essential aims in seat surface selection in order to promote a good microclimate. Development of support surface technologies relative to the seated position that can enable users to control heat, cooling and ventilation are of vital importance for users with autonomic dysfunction.

This paper will present a case study review of a wheelchair user with high level SCI, MS or CP and analyse their specific temperature dysregulation challenges. It will then focus on the use of a user controlled temperature controlled J3 back support, via a product development trial, to aid heat regulation in order to improve comfort and function for the user. The back support features an integral heat and ventilation pad within the foam overlay which includes fans and lithium battery. The user will use this equipment for a period of one month, and act as their own control by using the equivalent standard J3 backrest for one month. The user will report back on functional differences and their subjective rating of comfort over that time.

According to Davis et al (cited in Pinter et al, 2015) heat generation from physical exercise and increased ambient temperature are the most important heat challenges for MS sufferers. Heat sensitivity is common in MS and the accompanying fatigue leads to decreased motor function (Davis et al, 2010). Deficits in function and comfort caused by increased temperature may be reduced by removing heat stressors and promoting cooling.

According to Hagen (2015) temperature dysregulation is a commonly reported problem in individuals with high SCI, and excessive sweating is common. Guidance from the National Spinal Injuries Centre, Stoke Mandeville Hospital (2010) advises that strategies to promote temperature regulation are important to prevent situations of hypothermia or over-heating.

Discussion

The case study review will discuss the outcomes of the product development trial in relation to a wheelchair user with SCI, MS or CP.

References

- Davis, SL et al (2010). *Thermoregulation in multiple sclerosis*. Journal of applied Physiology. 109 (5) 1531-1537
- Hagen, E (2015). *Acute complications of spinal cord injuries*. World J Orthop 18, 6 (1), 17-23
- National Spinal Injuries Centre, Stoke Mandeville Hospital (2010). *Handout for Patients, Carers and Community Staff*.
- Pinter, A et al (2015). *Autonomic Dysregulation in Multiple Sclerosis* Int. J. Mol. Sci 16, 16920-16952 www.mdpi.com/journals/ijms

Email: matthew.eveleigh@sunmed.co.uk